


DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
6.11.20	All parts of Autoclave are checked		A	working
13.11.20	Gasket washed with distilled water		A	done
20.11.20	checked the compressed air line for any leakage		A	working
24.11.20	Nuts & bolts are checked		A	working
4.12.20	Heating coil changed		A	done
11.12.20	All parts of Autoclave are checked		A	working
12.12.20	Gasket washed with distilled water		A	done
26.12.20	Nuts & bolts are checked		A	working
02.01.21	Autoclave parts checked		A	working
9.01.21	Changed water weekly		A	done
18.01.21	Gasket washed with distilled water		A	done
25.1.21	Nuts & bolts are checked		A	working
1.2.21	Heating coil plug checked		A	working
8.2.21	All parts of Autoclave are checked		A	working
15.2.21	checked the compressed air line for any leakage		A	working
22.2.21	line for any leakage		A	done
1.3.21	Nuts & bolts are checked		A	working
8.3.21	Autoclave Maintenance checked		A	working
15.3.21	Nuts & bolts are checked		A	working
22.3.21	Heating coil checked		A	working
6.4.21	Distilled water Changed		A	done
15.4.21	checked the compressed air		A	working
22.4.21	line for any leakage		A	done
14.6.21	All the parts of the Autoclave		A	done
16.6.21	gasket washed thoroughly		A	done
24.6.21	Gasket washed with distilled water		A	done
01.07.21	Nuts & bolts are checked		A	working
08.07.21	Autoclave maintenance checked		A	working
15.7.21	indicator checked		A	working
22.7.21	Nuts & bolts checked		A	working
29.7.21	checked the compressed air line for any leakage		A	working



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
5.8.21	Gasket washed with soap distilled water	8	A	Working
12.8.21	Heating coil checked	8	A	Working
19.8.21	Indicator checked	8	A	Working
26.8.21	Nuts & bolts are checked	8	A	Working
2.9.21	Autoclave maintenance checked	8	A	Working
9.9.21	Changed the distilled water	8	A	Working
16.9.21	All the parts are checked	8	A	Working
23.9.21	Heating coil checked	8	A	Working
30.9.21	Gasket washed with distilled water	8	A	Working
7.10.21	All the parts of the autoclave	8	A	Working
18.10.21	are oiled & washed thoroughly	8	A	Working
25.10.21	Indicator was checked	8	A	Working
1.11.21	Heating coil plug checked	8	A	Working
8.11.21	Gasket changed	8	A	Working
15.11.21	Autoclave maintenance checked	8	A	Working
22.11.21	Maintenance done	8	A	Working
29.11.21	Nuts & bolts are checked	8	A	Working
6.12.21	Dust cleaned	8	A	Working
13.12.21	Gasket checked	8	A	Working
20.12.21	All the parts are checked	8	A	Working
27.12.21	Water changed	8	A	Working
3.1.22	Dust cleaned	8	A	Working
10.1.22	Heating coil checked	8	A	Working
17.1.22	Gasket washed with soap & distilled water	8	A	Working
24.1.22	Indicator checked	8	A	Working
31.1.22	Maintenance done	8	A	Working
7.2.22	Dust cleaned	8	A	Working
14.2.22	water changed	8	A	Working
21.2.22	Nuts & bolts are checked	8	A	Working
28.2.22	Checked the compressed air line	8	A	Working
7.3.22	for any leakage	8	A	Working
14.3.22	Heating coil checked	8	A	Working
21.3.22	Gasket washed with distilled water	8	A	Working



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
28.3.22	Maintenance done			
4.4.22	Dust checked the compressed air line for any leakage			working
11.4.22	Nuts and bolts are checked			working
13.4.22	Changed the distilled water			working
25.4.22	Heating coil checked			working
2.5.22	Dust cleaned			working
9.5.22	Gasket washed with distilled water			working
16.5.22	All the parts checked			working
23.5.22	Heating coil checked			working
30.5.22	Maintenance done			working
6.6.22	Dust cleaned			working
17.6.22	Nuts and bolts are checked			working
20.6.22	Indicator checked			working
27.6.22	Gasket washed with distilled water			working
4.7.22	Maintenance done			working
11.7.22	Dust cleaned			done
18.7.22	Checked nuts & bolts			working
25.7.22	Checked the compressed air line for any leakage			working
2/8/22	Dust cleaned			done
10/8/22	Nuts and bolts are checked			working
17/8/22	Indicator checked			working
26/8/22	checked the compressed air line for any leakage			working
3/9/22	Gasket washed with distilled water			done
2/9/22	Nuts and bolts are checked			working
14/9/22	Indicator checked			working
23/9/22	Maintenance done			done
1/10/22	Dust cleaned			done
11/10/22	Nuts & bolts are checked			working
20/10/22	Gasket washed with distilled water			working



Aditya Pharmacy College
SURAMPALAM-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
28/10/22	Heating coil checked			working
4/11/22	Indicator checked			working
10/11/22	Nuts & bolts are checked			working
16/11/22	Maintenance done			done
27/11/22	Changed the distilled water			done
2/12/22	Heating coil checked			working
1/12/22	Nuts & bolts are checked			working
6/12/22	Indicator checked			working
6.12.22	Nuts & bolts are checked			working
2.1.23	Heating coil checked			working
7.1.23	Maintenance was done			done
8.1.23	Changed the distilled water			done
8.1.23	Nuts & Bolts are checked			working
1.2.23	Heating coil checked			working
7.2.23	Maintenance done			done
6.2.23	Indicator was checked			working
22.2.23	Nuts & Bolts are checked			working
1.3.23	Heating coil was checked			working
1.3.23	Maintenance done			done
1.3.23	Nuts & Bolts are checked			done
1.3.23	Heating coil was checked			working
1.3.23	Maintenance done			done
8.4.23	Indicator was checked			working
5.4.23	Nuts & Bolts are checked			working
4.4.23	Maintenance done			done
5.23	changed the distilled water			done
5.23	Heating coil checked			working
5.23	Indicator checked			working
5.23	Nuts & Bolts are checked			working
7.5.23	Maintenance done			done
1.6.23	changed the distilled water			done
1.6.23	Heating coil checked			working
1.6.23	Indicator checked			working



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB PASS	SIGN OF LAB INCIDENTAL	REMARKS
6.11.20	Exterior & interior parts cleaned	8	A	working
7.11.20	Lubricant applied to the bucket grooves	8	A	working
13.11.20	checked the residue & corrosion on the rotors	8	A	working
22.11.20	knobs checked	8	A	working
1.12.20	Maintenance done	8	A	working
9.12.20	Exterior & interior parts cleaned	8	A	working
15.12.20	Lubricant applied to the bucket grooves	8	A	working
23.12.20	Checked the residue & corrosion on the rotors	8	A	working
2.1.21	knobs checked	8	A	working
11.1.21	Maintenance done	8	A	working
20.1.21	Exterior & interior parts cleaned	8	A	working
27.1.21	Lubricant applied to the bucket grooves	8	A	working
3.2.21	Residue & corrosion on the rotors checked	8	A	working
11.2.21	knobs checked	8	A	working
17.2.21	Maintenance done	8	A	working
23.2.21	Exterior & interior parts cleaned	8	A	working
2.3.21	Lubricant applied to the bucket grooves	8	A	working
8.3.21	Residue & corrosion on the rotors checked	8	A	working
16.05.21	knobs checked	8	A	working
23.3.21	Maintenance done	8	A	working
1.4.21	Exterior & interior parts cleaned	8	A	working
8.4.21	Lubricant applied to the bucket grooves	8	A	working
16.4.21	Residue & corrosion on the rotors checked	8	A	working
24.4.21	knobs checked	8	A	working
4.5.21	Maintenance done	8	A	working
20.5.21	Lubricant & corrosion on the rotors checked	8	A	working
3.06.21	Residue & corrosion on the rotors checked	8	A	working
11.6.21	Exterior & interior parts cleaned	8	A	working
21.06.21	knobs checked	8	A	working
1.7.21	Maintenance done	8	A	working
9.7.21	Lubricant applied to the bucket grooves	8	A	working



Aditya Pharmacy College
SURAMPAL-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
15-7-21	Exterior Interior parts cleaned	⌘	A	working
20-7-21	Lubricant applied to the bucket grooves	⌘	A	working
2-8-21	checked the residue & corrosion on rotors	⌘	A	working
10-8-21	knobs checked	⌘	A	working
16-8-21	Maintenance done	⌘	A	working
22-8-21	Exterior interior parts cleaned	⌘	A	working
3-9-21	Lubricant applied to the bucket grooves	⌘	A	working
9-9-21	checked the residue & corrosion on rotors	⌘	A	working
18-9-21	Knobs checked	⌘	A	working
28-9-21	Maintenance done	⌘	A	working
5-10-21	Exterior interior parts cleaned	⌘	A	working
12-10-21	Lubricant applied to the bucket grooves	⌘	A	working
20-10-21	checked the residual corrosion on rotors	⌘	A	working
28-10-21	knobs checked	⌘	A	working
28-10-21	Maintenance done	⌘	A	working
6-11-21	Exterior interior parts cleaned	⌘	A	working
15-11-21	Lubricant applied to the bucket grooves	⌘	A	working
23-11-21	checked the residue & corrosion on rotors	⌘	A	working
30-11-21	knobs checked	⌘	A	working
8-12-21	Maintenance done	⌘	A	working
16-12-21	Exterior & interior parts cleaned	⌘	A	working
24-12-21	Lubricant applied to the bucket grooves	⌘	A	working
30-12-21	checked the residue & corrosion on rotors	⌘	A	working
3-01-22	knobs checked	⌘	A	working
12-01-22	Maintenance done	⌘	A	working
20-1-22	Exterior & interior parts cleaned	⌘	A	working
28-01-22	Lubricant applied to the bucket grooves	⌘	A	working
4-02-22	checked the residue & corrosion on rotors	⌘	A	working
12-2-22	knobs checked	⌘	A	working
20-2-22	Maintenance done	⌘	A	working
28-2-22	Exterior & interior parts cleaned	⌘	A	working
2-3-22	Lubricant applied to the bucket grooves	⌘	A	working
10-3-22	checked the residue & corrosion rotors	⌘	A	working



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
19.3.22	Knobs checked	8	A	
28.3.22	Maintenance done	8	A	working
1.4.22	Exterior & interior parts cleaned	8	A	working
11.4.22	Lubricant applied to the bucket grooves	8	A	working
18.4.22	Checked the residue & corrosion on motor	8	A	working
25.4.22	Knobs checked	8	A	working
2.5.22	Maintenance done	8	A	working
10.5.22	Lubricant applied to the bucket grooves	8	A	working
19.5.22	Checked the residue & corrosion on motor	8	A	working
27.5.22	Knobs checked	8	A	working
4.6.22	Maintenance done	8	A	working
11.6.22	Exterior & interior parts cleaned	8	A	working
20.6.22	Lubricant applied to the bucket grooves	8	A	working
28.6.22	Knobs checked	8	A	working
1.7.22	Dust cleaned	8	A	working
18.7.22	Exterior & interior parts cleaned	8	A	done
22.7.22	Lubricant applied to the bucket grooves	8	A	done
29.7.22	Maintenance was done	8	A	working
3/8/22	Dust cleaned	8	A	done
10/8/22	Knobs checked	8	A	working
16/8/22	Exterior & interior parts cleaned	8	A	done
24/8/22	Lubricant app. Maintenance done	8	A	done
2/9/22	Dust cleaned	8	A	done
9/9/22	Knobs checked	8	A	working
15/9/22	Exterior & interior parts cleaned	8	A	done
25/9/22	Maintenance done	8	A	done
1/10/22	Dust cleaned	8	A	done
10/10/22	Knobs checked	8	A	working
18/10/22	Exterior & interior parts cleaned	8	A	done
26/10/22	checked the residue & corrosion on motor	8	A	working
4/11/22	Dust cleaned	8	A	done
12/11/22	Knobs checked	8	A	working
19/11/22	Lubricant applied to the bucket grooves	8	A	done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

30/11/22	Exterior & Interior parts cleaned	g	g	done
16.12.22	Lubricant applied to bucket gears	g	g	done
26.12.22	knobs checked	g	g	working
9.1.23	Maintenance done	g	g	done
18.1.23	checked all electrical terminals	g	g	working
2.2.23	cleaning with wet mopping	Adh	g	done
9.2.23	knobs are checked	Adh	g	working
23.2.23	Maintenance was done	Adh	g	done
10.3.23	checked all electrical terminals	Adh	g	working
17.3.23	cleaning with wet mopping	Adh	g	done
24.3.23	knobs : are checked	Adh	g	work
31.3.23	Maintenance was done	Adh	g	done
15.4.23	checked all electrical terminals	Adh	g	working
24.4.23	cleaning with wet mopping	Adh	g	done
1.5.23	knobs are checked	Adh	g	working



PRINCIPAL

Aditya Pharmacy College
SURAMPALAM-533 437

NAME OF THE EQUIPMENT : COMPOUND MICROSCOPE

EQUIPMENT CODE : APC-P. Log-SOP-201/202-07

1

DATE	EQUIPMENT MAINTENANCE	Sign of LAB Tech	Sign of LAB INCHARGE	REMARKS
7.11.20	Eye piece & objective lens cleaned	8	A	done
16.11.20	Moment of knobs was checked	8	A	working
23.11.20	Lubricant used to the stage	8	A	done
30.11.20	Condenser was checked	8	A	working
7.12.20	Eye piece & objective lens cleaned	8	A	done
14.12.20	Moment of knobs was checked	8	A	working
21.12.20	Lubricant used to the stage	8	A	done
29.12.20	Condenser & mirror checked	8	A	working
4.1.21	Maintenance done	8	A	done
11.1.21	Eye piece & objective lens cleaned	8	A	done
18.1.21	Moment of knobs was checked	8	A	working
25.1.21	Lubricant used to the stage	8	A	done
2.2.21	Condenser & mirror was checked	8	A	working
9.2.21	Maintenance done	8	A	done
16.2.21	Eye piece & objective lens cleaned	8	A	done
23.2.21	Moment of knobs was checked	8	A	working
2.2.21	Lubricant used to the stage	8	A	done
9.3.21	Condenser & mirror checked	8	A	working
16.3.21	Maintenance done	8	A	done
23.3.21	Eye piece & objective lens cleaned	8	A	done
30.3.21	Lubricant used to the stage	8	A	working
7.4.21	Moment of knobs was checked	8	A	working
15.4.21	Condenser & mirror checked	8	A	working
22.4.21	Maintenance done	8	A	done
29.4.21	Eye piece & objective lens cleaned	8	A	done
11.6.21	Lubricant used to the stage	8	A	working
18.6.21	Condenser & mirror checked	8	A	working
25.6.21	Moment of knobs was checked	8	A	working
2.7.21	Maintenance done	8	A	done
9.7.21	Eye piece & objective lens cleaned	8	A	done
16.7.21	Lubricant used to the stage	8	A	working
23.7.21	Condenser & mirror checked	8	A	done
30.7.21	Maintenance done	8	A	done



PRINCIPAL
Aditya Pharmacy College
SURAMPATEM-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
3.8.21	Eyepiece & objective lens cleaned	8	A	working
9.8.21	Moment of knobs checked	8	A	working
17.8.21	Lubricant used to the stage	8	A	working
23.8.21	Condenser & mirror checked	8	A	working
4.9.21	Eyepiece & objective lens cleaned	8	A	working
8.9.21	Moment of knobs checked	8	A	working
5.9.21	Lubricant used to the stage	8	A	working
14.9.21	Condenser & mirror checked	8	A	working
1.10.21	Maintenance done	8	A	working
11.10.21	Eyepiece & objective lens cleaned	8	A	working
8.10.21	Moment of knobs checked	8	A	working
17.10.21	Lubricant used to the stage	8	A	working
2.11.21	Condenser & mirror checked	8	A	working
8.11.21	Eyepiece & objective lens cleaned	8	A	working
6.11.21	Moment of knobs checked	8	A	working
13.11.21	Lubricant used to the stage	8	A	working
3.12.21	Maintenance done	8	A	working
1.12.21	Eyepiece & objective lens cleaned	8	A	working
5.12.21	Moment of knobs checked	8	A	working
4.12.21	Lubricant used to the stage	8	A	working
1.1.22	condenser & mirror checked	8	A	working
3.1.21	Eyepiece & objective lens cleaned	8	A	working
7.1.21	moment of knobs checked	8	A	working
15.1.21	lubricant used to the stage	8	A	working
1.2.21	Condenser & mirror checked	8	A	working
2.2.21	Eyepiece & objective lens cleaned	8	A	working
20.2.21	Moment of knobs checked	8	A	working
28.2.21	Lubricant applied to the stage	8	A	working
2.3.21	Condenser & mirror checked	8	A	working
10.3.21	Eyepiece & objective lens cleaned	8	A	working
19.3.21	Moment of knobs checked	8	A	working
28.3.21	Lubricant applied to the stage	8	A	working
1.4.21	condenser & mirror checked	8	A	working



PRINCIPAL
Aditya Pharmacy College
Srirampalem-533 437

DATE	EQUIPMENT MAINTENANCE	SIGN OF LAB ASST	SIGN OF LAB INCHARGE	REMARKS
11.4.22	Eye piece & objective lens cleaned	8	A	working
17.4.22	Movement of knobs checked	8	A	working
25.4.22	Lubricant used to the stage	8	A	working
2.5.22	Condenser & mirror checked	8	A	working
10.5.22	Maintenance done	8	A	working
19.5.22	Eye piece & objective lens cleaned	8	A	working
27.5.22	Movement of knobs checked	8	A	working
4.6.22	Lubricant used to the stage	8	A	working
12.6.22	Condenser & mirror checked	8	A	working
20.6.22	Maintenance done	8	A	working
28.6.22	Eye piece & objective lens cleaned	8	A	working
7.7.22	Movement of knobs checked	8	A	working
14.7.22	Lubricant used to the stage	8	A	done
20.7.22	Condenser & mirror checked	8	A	working
27.7.22	Maintenance was done	8	A	done
3/8/22	Eye piece & objective lens cleaned	8	A	done
11/8/22	Movement of knobs checked	8	A	working
16/8/22	Lubricant used to the stage	8	A	working
29/8/22	Condenser & mirror checked	8	A	working
2/9/22	Maintenance done	8	A	done
9/9/22	Eye piece & objective lens cleaned	8	A	done
15/9/22	Movement of knobs checked	8	A	working
25/9/22	Lubricant used to the stage	8	A	done
1/10/22	Condenser & mirror checked	8	A	working
10/10/22	Maintenance done	8	A	done
17/10/22	Eye piece & objective lens cleaned	8	A	done
26/10/22	Condenser & mirror checked	8	A	working
4/11/22	Movement of knobs checked	8	A	working
12/11/22	Lubricant used to the stage	8	A	done
19/11/22	Maintenance done	8	A	done
30/11/22	Eye piece & objective lens cleaned	8	A	working
2/12/22	Movement of knobs checked	8	A	done
9/12/22	Lubricant used to the stage	8	A	done



PRINCIPAL

Aditya Pharmacy College
SURAMPALAM 533 437

1/12/22	Condensor & mirror checked	g	g	working
2/12/22	Maintenance was done	g	g	done
7.1.23	Eye piece and objective lens cleaned	g	g	done
8.1.23	Knobs movement checked	g	g	working
2.2.23	Lubricant was applied to stage	Appl	g	done
7.2.23	Maintenance was done	Appl	g	done
3.2.23	Eye piece and objective lens cleaned	Appl	g	done
4.3.23	Knobs movement checked	Appl	g	working
1.3.23	Lubricant used to the stage	Appl	g	done
1.2.23	Maintenance was done	Appl	g	done
1.3.23	Eye piece and objective lens cleaned	Appl	g	done
1.3.23	Knobs movement checked	Appl	g	working
1.4.23	Lubricant used to the stage	Appl	g	done
1.4.23	Maintenance was done	Appl	g	done
✓ = 7		Ad		1




 PRINCIPAL
 Aditya Pharmacy College
 SURAMPALAM-533 437

Date	Equipment Maintenance	Sign of the lab/	Sign of the lab/	Remark
21-03-22	checked the screws fitting of the Autoclave	T BSH	Ceth	Done
28-03-22	Maintenance done	T BSH	Ceth	Done
14-04-22	cleaned the equipment with the detergent	T BSH	Ceth	Done
1-04-22	Dust cleaned	T BSH	Ceth	Done.
8-04-22	checked the screws fitting of the Autoclave	T BSH	Ceth	Working
5-04-22	sterilize the instrument with distilled water and dried it checked for any leaks of equipment	T BSH	Ceth	Working
12-05-22	cleaned the equipment with the detergent	T BSH	Ceth	Done
9-05-22	checked the screws fitting of the Autoclave	T BSH	Ceth	Done
6-05-22	Dust was cleaned	T BSH	Ceth	Done
13-05-22	Maintenance was done	T BSH	Ceth	Done
30-05-22	checked the screws fitting of the Autoclave	T BSH	Ceth	Working
6-06-22	cleaned the equipment with the detergent	T BSH	Ceth	Done
7-06-22	Maintenance was done	T BSH	Ceth	Done
0-06-22	Dust was cleaned	T BSH	Ceth	Done
17-06-22	All parts checked & cleaned	T BSH	Ceth	Working
14-07-22	Maintenance was done	T BSH	Ceth	Done
12-07-22	checked the screws fitting of the Autoclave	T BSH	Ceth	Done
11-07-22	Maintenance was done	T BSH	Ceth	Done
9-07-22	cleaned the equipment with the detergent.	T BSH	Ceth	Done
12-08-22	Dust cleaned	T BSH	Ceth	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-5200137

Date	Equipment Maintenance	Sign of the lab/T	Sign of the lab/E	Status
10-08-22	checked for screws fitting of the Autoclave	T. PAI	PAI	Working
18-08-22	Maintenance done	T. PAI	PAI	Done
24-08-22	cleaned the equipment with detergent	T. PAI	PAI	Done
03-09-22	Dust cleaned.	T. PAI	PAI	Done
12-09-22	checked the Instrument for any leakage	T. PAI	PAI	Done
20-09-22	Sterilized the Instrument with distilled water	T. PAI	PAI	Done
28-09-22	Dust cleaned.	T. PAI	PAI	Done
01-10-22	cleaned the equipment with detergent	T. PAI	PAI	Done
11-10-22	checked the screws fitting of Autoclave.	T. PAI	PAI	Working
19-10-22	Maintenance done	T. PAI	PAI	Done
27-10-22	All parts checked & cleaned.	T. PAI	PAI	Working
04-11-22	checked any leaks of equipment	T. PAI	PAI	Done
12-11-22	Dust cleaned.	T. PAI	PAI	Done
19-11-22	cleaned the equipment with detergent	T. PAI	PAI	Done
26-11-22	Maintenance done	T. PAI	PAI	Done
03-12-22	Sterilized the Instrument with distilled water	T. PAI	PAI	Done
11-12-22	checked the Instrument for any leakage	T. PAI	PAI	Working
20-12-22	Dust cleaned	T. PAI	PAI	Done
31-12-22	All parts checked and cleaned	T. PAI	PAI	Done
6-1-23	checked the screws fitting of Autoclave	T. PAI	PAI	Done
20-1-23	Maintenance was done.	T. PAI	PAI	Done
27-1-23	checked any leaks of equipment	T. PAI	PAI	Done
3-2-23	Dust was cleaned	T. PAI	PAI	Done
10-2-23	cleaned the equipment with detergent	T. PAI	PAI	Done
17-2-23	Sterilized the Instrument with distilled water	T. PAI	PAI	Done
24-2-23	checked the screws fitting of Autoclave.	T. PAI	PAI	Working
3-3-23	Maintenance was done.	T. PAI	PAI	Done
10-3-23	checked any leaks of equipment	T. PAI	PAI	Done
17-3-23	All parts checked and cleaned.	T. PAI	PAI	Working
24-3-23	cleaned the equipment with detergent	T. PAI	PAI	Done
31-3-23	checked the Instrument for any leakage	T. PAI	PAI	Done
3-4-23	Dust was cleaned	T. PAI	PAI	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

ADITYA PHARMACY COLLEGE

Date	Equipment Maintenance	sign of the lab	sign of the lab	Person
26-11-20	Cleaned the eyepiece, mirror	T. Test	C. Test	Done
13-11-20	Condensor with lens cleaning fluid.	T. Test	C. Test	Done
20-11-20	Wiped the immersion oil of the objective, Condensor & Mechanical stage with muslin cloth.	T. Test	C. Test	Done
17-11-20	checked for the magnification of lenses	T. Test	C. Test	Done
14-12-20	Maintenance done	T. Test	C. Test	Done
18-12-20	cleaned the eye piece, mirror.	T. Test	C. Test	Done
31-12-20	condensor with lens cleaning fluid	T. Test	C. Test	Done
	Wiped the immersion of the objective, condensor & Mechanical stage with muslin cloth			
02-01-21	checked for the magnification of lenses.	T. Test	C. Test	Done
09-01-21	Maintenance done	T. Test	C. Test	Done
13-01-21	cleaned the eyepiece, mirror, condensor with lens cleaning fluid	T. Test	C. Test	Done
25-01-21	Maintenance done	T. Test	C. Test	Done
01-02-21	checked for the magnification of lenses	T. Test	C. Test	Done
08-02-21	cleaned the eyepiece, mirror.	T. Test	C. Test	Done
15-02-21	condensor with lens fluid.	T. Test	C. Test	Done
	wiped the immersion oil of the objective, condensor, Mechanical stage with muslin cloth.			
22-02-21	maintenance done			Done
01-03-21	Lubricant used on the Stage	T. Test	C. Test	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment Maintenance	Sign of the lab/T	Sign of the lab/T	Remark
08-03-21	cleaned the eyepiece, mirror	T Test	Cyru	Done
15-03-21	condensor with lens containing fluid	T Test	Cyru	Done
22-03-21	checked for the magnification of lenses	T Test	Cyru	Done
06-04-21	Lubricant used to the stage	T Test	Cyru	Done
15-04-21	cleaned the eyepiece mirror	T Test	Cyru	Done
22-04-21	condensor with lens containing	T Test	Cyru	Done
10-06-21	maintenance done	T Test	Cyru	Done
16-06-21	Lubricant used to the stage	T Test	Cyru	Done
24-06-21	checked for the magnification of lenses	T Test	Cyru	Done
01-07-21	cleaned the eye piece, mirror	T Test	Cyru	Done
08-07-21	condensor with lens containing fluid	T Test	Cyru	Done
15-07-21	Maintenance done	T Test	Cyru	
22-07-21	Lubricant used to the stage	T Test	Cyru	Working
29-07-21	checked for the magnification of lenses	T Test	Cyru	Done
05-08-21	cleaned the eye piece, mirror condensor with lens cleaning fluid	T Test	Cyru	Done
12-08-21	checked for the magnification of lenses	T Test	Cyru	Done
19-08-21	Lubricant used to the stage	T Test	Cyru	Working
26-08-21	Maintenance done	T Test	Cyru	
2-09-21	cleaned the eye piece, mirror, condensor with lens cleaning fluid	T Test	Cyru	Done



Aditya Pharmacy College
SURAMPALEM-533 437

Date	Equipment Maintenance	Sign of the lab T	Sign of the lab R	Remark
02-09-21	Equipment light source was checked	T Test	Cygn	Done
09-09-21	Iris diaphragm was checked	T Test	Cygn	Working
16-09-21	Ocular focus was checked	T Test	Cygn	Working
23-09-21	Course adjustment was checked	T Test	Cygn	Done
30-09-21	Mechanical stage and clips was checked	T Test	Cygn	Working
17-10-21	Mechanical slide holder was checked	T Test	Cygn	Working
18-10-21	Iris diaphragm condenser was checked	T Test	Cygn	Done
25-10-21	Condenser adjustment knob was checked	T Test	Cygn	Done
01-11-21	Fine adjustment knob was checked	T Test	Cygn	Done
08-11-21	Cleaned lenses with tissue paper	T Test	Cygn	Done
15-11-21	Dust was cleaned	T Test	Cygn	Done
22-11-21	Maintenance was done	T Test	Cygn	Done
29-11-21	Applied emulsions to the knobs and screws	T Test	Cygn	Done
06-12-21	checked the equipment plug and wire	T Test	Cygn	Working
13-12-21	Equipment light source was checked	T Test	Cygn	Done
20-12-21	Iris diaphragm was checked	T Test	Cygn	Working
27-12-21	Ocular focus was checked	T Test	Cygn	Done
03-01-22	Course adjustment was checked	T Test	Cygn	Done
10-01-22	Fine adjustment knob was checked	T Test	Cygn	Done
17-01-22	Cleaned lenses with tissue paper	T Test	Cygn	Done
24-01-22	Dust was cleaned	T Test	Cygn	Done
30-01-22	Maintenance was done	T Test	Cygn	Done
07-02-22	Applied emulsion to the knobs and screws	T Test	Cygn	Done
	checked the equipment plug	T Test	Cygn	Working



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment Maintenance	sign of the lab	sign of the lab	Remarks
21-02-22	Equipment light source was checked	T Datt	Cypr	Done
28-02-22	Iris diaphragm was checked	T Datt	Cypr	Working
14-03-22	ocular focus was checked	T Datt	Cypr	Done
23-03-22	course adjustment was checked	T Datt	Cypr	Done
29-03-22	Mechanical stage and clips were checked	T Datt	Cypr	Done
06-04-22	Mechanical slide holder was checked	T Datt	Cypr	
14-04-22	Iris diaphragm condenser was checked condenser adjustment knob was checked	T Datt	Cypr	Working Working
21-04-22	Fine adjustment knob was checked	T Datt	Cypr	Done
28-04-22	Cleaned lenses with tissue Paper.	T Datt	Cypr	Done
06-05-22	Dust was cleaned	T Datt	Cypr	Done
12-05-22	Maintenance was done	T Datt	Cypr	Done
	Applied emulsion to the knobs and screws	T Datt	Cypr	Done
18-06-22	checked the equipment plug and wire	T Datt	Cypr	Working
12-08-22	Equipment light source was checked	T Datt	Cypr	Done
19-08-22	Iris diaphragm was checked	T Datt	Cypr	Working
	ocular focus was checked	T Datt	Cypr	Done
	course adjustment was checked	T Datt	Cypr	Done
26-08-22	Mechanical stage and clips was checked	T Datt	Cypr	Done
	Mechanical slide holder was checked	T Datt	Cypr	Done
02-09-22	Iris diaphragm condenser checked	T Datt	Cypr	Working
09-09-22	condenser adjustment knob was checked	T Datt	Cypr	Done



PRINCIPAL


Aditya Pharmacy College
SURAMPALAM 533 437

Date	Equipment Maintenance	Sign of the bbl/r	Sign of the bbl/r	Remark
14-9-22	Fine adjustment knob was checked	T Bell	Cyru	Done
	cleaned lenses with tissue paper	T Bell	Cyru	Done
27-9-22	Dust was cleaned	T Bell	Cyru	Done
30-9-22	Maintenance was done	T Bell	Cyru	Done
07-10-22	Applied emulsion to the knobs and screws	T Bell	Cyru	Done
18-10-22	checked the equipment plug	T Bell	Cyru	working
	cleaned the eyepiece, mirror condenser with lens fluid	T Bell	Cyru	Done
25-10-22	Equipment light source was checked	T Bell	Cyru	Done
01-11-22	Iris diaphragm was checked	T Bell	Cyru	working
08-11-22	Ocular focus was checked	T Bell	Cyru	Done
15-11-22	coarse adjustment was checked	T Bell	Cyru	Done
22-11-22	Mechanical stage and clips was checked	T Bell	Cyru	Done
06-12-22	Mechanical slide holder was checked	T Bell	Cyru	Done
13-12-22	Iris diaphragm condenser was checked	T Bell	Cyru	working
20-12-22	condenser adjustment knob was checked	T Bell	Cyru	Done
27-12-22	Fine adjustment knob was checked	T Bell	Cyru	Done
	cleaned lenses with tissue paper	T Bell	Cyru	Done
03-01-23	Dust was cleaned	T Bell	Cyru	Done
10-01-23	Maintenance was done	T Bell	Cyru	Done
17-01-23	Applied emulsion to the knobs and screws	T Bell	Cyru	Done
24-01-23	checked the equipment plug and wire	T Bell	Cyru	working
30-01-23	Dust was cleaned	T Bell	Cyru	Done
07-02-23	cleaned the eyepiece, mirror condenser with lens fluid	T Bell	Cyru	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALEM-533 437

Date	Equipment Maintenance	Sign of the lab/T	Sign of the lab/E	Remark
14-03-22	checked the plug and wire	T. Deth	Lab	Working
21-03-22	fine adjustment knob was checked	T. Deth	Lab	Working
28-03-22	Maintenance done	T. Deth	Lab	Done
04-04-22	checked for magnification of lense.	T. Deth	Lab	Working
11-04-22	Lubricant used to the stage	T. Deth	Lab	Done.
18-04-22	Dust- cleaned.	T. Deth	Lab	Done.
25-04-22	cleaned the eyepiece, mirror Condensed the lens with fluid	T. Deth	Lab	Done.
02-05-22	fine adjustment knob was checked.	T. Deth	Lab	Working
09-05-22	Equipment light source was checked.	T. Deth	Lab	Working
16-05-22	Mechanical stage was checked.	T. Deth	Lab	Working
23-05-22	Dust cleaned.	T. Deth	Lab	Done
30-05-22	cleaned lenses with tissue paper	T. Deth	Lab	Done
06-06-22	Condens- adjustment knob was checked.	T. Deth	Lab	Working
17-06-22	Maintenance done.	T. Deth	Lab	Done.
20-06-22	checked the equipment plug and wire	T. Deth	Lab	Working
27-06-22	Irish diaphragm checked	T. Deth	Lab	Working
04-07-22	Condens- adjustment knob was checked.	T. Deth	Lab	Working
12-07-22	dust was cleaned.	T. Deth	Lab	Done




PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment Maintenance	Sign of the lab	Sign of the lab	Remark
20-07-22	Maintenance was done.	T. Bell	Ch	Done
25-07-22	Lubricant used to the Stage.	T. Bell	Ch	Done
01-08-22	Fine adjustment knob was checked.	T. Bell	Ch	Working
08-08-22	Dust cleaned.	T. Bell	Ch	Done
16-08-22	Equipment light source was checked.	T. Bell	Ch	Working
22-08-22	Dust cleaned		Ch	
01-09-22	iris diaphragm Condenser was checked.	T. Bell	Ch	Working
09-09-22	Mechanical Stage and clips was checked.	T. Bell	Ch	Working
17-09-22	Fine adjustment knob was checked.	T. Bell	Ch	Working
26-09-22	Lubricant used to the Stage	T. Bell	Ch	Done
10-10-22	Maintenance done.	T. Bell	Ch	Done
17-10-22	cleaned lenses with tissue paper.	T. Bell	Ch	Done
24-10-22	Condenser adjustment knob was checked.	T. Bell	Ch	Working
29-10-22	cleaned the eyepiece, iris diaphragm Condenser the lens with fluid.	T. Bell	Ch	Done
04-11-22	Dust cleaned.	T. Bell	Ch	Done
12-11-22	checked the equipment plug and wire.	T. Bell	Ch	Working
18-11-22	Maintenance done.	T. Bell	Ch	Done
29-11-22	Dust cleaned.	T. Bell	Ch	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment maintenance	Sign of the blb/	Sign of the blb/	Remark
07-12-22	checked the plug and wire	T. Dill	Self up	working
14-12-22	Lubricant used to the stage	T. Dill	CB	Done
20-12-22	cleaned lenses with tissue paper	T. Dill	CB	Done
24-12-22	checked for magnification of lense	T. Dill	CB	working
31-12-22	fine adjustment knob was checked	T. Dill	CB	working
6-1-23	cleaned the lense with tissue paper	T. Dill	CB	Done
20-1-23	Maintenance was done	T. Dill	CB	Done
27-1-23	Condenser knob was checked	T. Dill	CB	working
3-2-23	lubricant used to the stage	T. Dill	CB	Done
10-2-23	checked the plug and wire	T. Dill	CB	working
17-2-23	cleaned the lense with tissue paper	T. Dill	CB	Done
24-2-23	Trich diaphragm was checked	T. Dill	CB	working
3-3-23	Dust was cleaned	T. Dill	CB	Done
10-3-23	lubricant used to the stage	T. Dill	CB	Done
17-3-23	fine adjustment knob was checked	T. Dill	CB	working
24-3-23	Maintenance was done	T. Dill	CB	Done
31-3-23	Condenser knob was checked	T. Dill	CB	working
3-4-23	Equipment light source was checked	T. Dill	CB	working
10-4-23	Trich diaphragm was checked	T. Dill	CB	working
17-4-23	cleaned the lense with tissue paper	T. Dill	CB	Done
24-4-23	lubricant used to the stage	T. Dill	CB	Done
1-5-23	Cleaned lense with tissue paper	T. Dill	CB	Done
4-5-23	fine adjustment knob was checked	T. Dill	CB	working
8-5-23	Condenser knob was checked	T. Dill	CB	working
15-5-23	Dust was cleaned	T. Dill	CB	Done
25-5-23	checked the plug and wire	T. Dill	CB	working
29-5-23	cleaned the lense with tissue paper	T. Dill	CB	Done
5-6-23	checked for the magnification of lense	S.	CB	Done
12-6-23	lubricant used to the stage	S.	CB	Done
19-6-23	cleaned lenses with tissue paper	S.	CB	Done



ADITYA PHARMACY COLLEGE
RAMPALEM-533 437

ADITYA PHARMACY COLLEGE

41

Date	Equipment Maintenance	Sign of the lab/1	Sign of the lab/2	Remark
07-11-20	checked all the electrical terminals of main supply	T. DBH	Cypr	working
16-11-20	checked heater connection	T. DBH	Cypr	working
28-11-20	checked cooling coils for any leakages	T. DBH	Cypr	Done
30-11-20	checked temperature controller & sensors for proper working	T. DBH	Cypr	Done
10-12-20	Dust was cleaned	T. DBH	Cypr	Done
18-12-20	Maintenance was done	T. DBH	Cypr	Done
25-12-20	checked all the electrical terminals of main supply	T. DBH	Cypr	working
02-01-21	checked heater connection	T. DBH	Cypr	Done
09-01-21	checked cooling coils for any leakage	T. DBH	Cypr	Done
18-01-21	checked temperature controller & sensors for proper working	T. DBH	Cypr	working
27-01-21	Dust was cleaned	T. DBH	Cypr	Done
02-02-21	Maintenance was Done	T. DBH	Cypr	Done
10-02-21	checked all the electrical terminal of main supply	T. DBH	Cypr	working
19-02-21	checked heater connection	T. DBH	Cypr	Done
26-02-21	checked cooling coils for any leakages	T. DBH	Cypr	Done
01-03-21	checked temperature controller & sensors for proper working	T. DBH	Cypr	working
10-03-21	Dust was cleaned	T. DBH	Cypr	Done
18-03-21	Maintenance was Done	T. DBH	Cypr	Done
27-03-21	checked all the electrical terminals of main supply	T. DBH	Cypr	working
23-04-21	checked heater connection	T. DBH	Cypr	Done
11-04-21	checked cooling coils for any	T. DBH	Cypr	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment Maintenance	sign of the lab/T	sign of the lab/T	Remarks
20-04-21	leakage checked temperature controller & sensors for proper working	T Deth	Cypr	working
28-04-21	Dust was cleaned	T Deth	Cypr	Done
04-05-21	Maintenance was done	T Deth	Cypr	Done
30-05-21	checked all the electrical terminals of main supply	T Deth	Cypr	working
08-06-21	checked heater connections	T Deth	Cypr	Done
16-06-21	checked cooling coils for any leakage	T Deth	Cypr	Done
24-06-21	checked temperature controller & sensors for proper working	T Deth	Cypr	working
30-06-21	Dust was cleaned	T Deth	Cypr	Done
02-07-21	Maintenance was done	T Deth	Cypr	Done
10-07-21	checked all the electrical terminals of main supply	T Deth	Cypr	working
19-07-21	checked heater connections	T Deth	Cypr	Done
26-07-21	checked cooling coils for any leakage	T Deth	Cypr	Done
02-08-21	checked temperature controller & sensors for proper working	T Deth	Cypr	working
10-08-21	Dust was cleaned	T Deth	Cypr	Done
19-08-21	checked all the electrical terminals of main supply	T Deth	Cypr	Done
27-08-21	checked heater conditions	T Deth	Cypr	Done
03-09-21	checked cooling coils for any leakage	T Deth	Cypr	Done
11-09-21	checked temperature controller & sensors for proper working	T Deth	Cypr	working
18-09-21	Dust was cleaned	T Deth	Cypr	Done
28-09-21	checked all the electrical	T Deth	Cypr	Done



Aditya Pharmacy College
SURAMPALAM-533 4111

Date	Equipment Maintenance	Sign of the lab/T	Sign of the lab/T	Remark
01-10-21	Terminals of main supply checked heater connection	T Datta	Cygn	working
12-10-21	Maintenance was Done	T Datta	Cygn	Done
20-10-21	checked cooling coils for any leakages	T Datta	Cygn	Done
28-10-21	checked temperature controller & sensors for proper working	T Datta	Cygn	working
02-11-21	Dust was cleaned	T Datta	Cygn	Done
11-11-21	Maintenance was done	T Datta	Cygn	Done
19-11-21	checked all the electrical terminals of main supply	T Datta	Cygn	working
26-11-21	checked heater connections	T Datta	Cygn	Done
02-12-21	checked cooling coils for any leakages	T Datta	Cygn	Done
10-12-21	checked temperature controller & sensors for proper working	T Datta	Cygn	working
20-12-21	Dust was cleaned	T Datta	Cygn	Done
28-12-21	checked all the electrical terminals of main supply	T Datta	Cygn	working
02-01-22	checked heater connections	T Datta	Cygn	Done
10-01-22	checked cooling coils for any leakages	T Datta	Cygn	Done
19-01-22	checked temperature controller & sensors for proper working	T Datta	Cygn	working
27-01-22	Maintenance was done	T Datta	Cygn	Done
03-02-22	checked all the electrical terminals of main supply	T Datta	Cygn	working
11-02-22	checked heater connections	T Datta	Cygn	Done
19-02-22	checked cooling coils for any leakages	T Datta	Cygn	Done
28-02-22	checked temperature controller & sensors for proper working	T Datta	Cygn	working



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment Maintenance	Sign of the lab/r	Sign of the lab/r	Remark
04-03-22	Dust was cleaned	T. Data	<u>Cyru</u>	Done
	checked all the electrical terminals of main supply	T. Data	<u>Cyru</u>	Working
12-03-22	checked heater connections	T. Data	<u>Cyru</u>	Done
	checked working coils for any leakage			Done
21-03-22	checked temperature controls & sensors for proper working	T. Data	<u>Cyru</u>	Working
	Dust was cleaned	T. Data		Done
26-03-22	checked all the electrical terminals of main supply	T. Data	<u>Cyru</u>	Working
01-04-22	checked heater connections	T. Data	<u>Cyru</u>	Done
	checked working coils for any leakages	T. Data	<u>Cyru</u>	Done
10-04-22	checked temperature controller & sensors for proper working	T. Data	<u>Cyru</u>	Working
	Maintenance was done	T. Data	<u>Cyru</u>	Done
19-04-22	checked all the electrical terminals of main supply	T. Data	<u>Cyru</u>	Working
	checked heater connections	T. Data	<u>Cyru</u>	Done
	checked working coils for any leakages	T. Data	<u>Cyru</u>	Done
27-04-22	checked temperature controller & sensor for proper working	T. Data	<u>Cyru</u>	Working
	Dust was cleaned	T. Data	<u>Cyru</u>	
02-05-22	Maintenance was done	T. Data	<u>Cyru</u>	Done
10-05-22	checked all the electrical terminals of main supply	T. Data	<u>Cyru</u>	Working
	checked heater connections	T. Data	<u>Cyru</u>	Done
	checked working coils for any leakage	T. Data	<u>Cyru</u>	Done
20-05-22	Dust was cleaned	T. Data	<u>Cyru</u>	Done



PRINCIPAL
Aditya Pharmacy College
SURAMPALAM-533 437

Date	Equipment Maintenance	Sign of the lab/1	Sign of the lab/2	Remark
29-05-22	checked temperature controller & sensors for proper working Maintenance was done	T. Dett	Cipm	Working Done
02-06-22	checked all the electrical terminal of main supply checked heater connection checked cooling coils for any leakage checked temperature controller & sensors for proper working	T. Dett T. Dett T. Dett T. Dett	Cipm Cipm Cipm Cipm	Working Working Done Done
10-06-22	Dust was cleaned Maintenance was done	T. Dett T. Dett	Cipm Cipm	Done Done
20-06-22	checked all the electrical terminal of main supply checked heater connection checked cooling coils for any leakage	T. Dett T. Dett T. Dett	Cipm Cipm Cipm	Working Done Done
27-06-22	checked temperature controller & sensors for proper working	T. Dett	Cipm	Working
01-07-22	Dust was cleaned checked temperature controller & sensors for proper working	T. Dett T. Dett	Cipm Cipm	Done Working
09-07-22	checked all the electrical terminal of main supply checked heater connection checked cooling coils for any leakage	T. Dett T. Dett T. Dett T. Dett	Cipm Cipm Cipm Cipm	Working Working Done Done
18-07-22	checked temperature controller & sensors for proper working	T. Dett	Cipm	Working
25-07-22	Dust was cleaned	T. Dett	Cipm	Done
02-08-22	Maintenance was done.	T. Dett	Cipm	Done



Principal
Aditya Pharmacy College
SURAMPALÉM-533 42

Date	Equipment maintenance	Slip of the table	Slip of the table	Status
14-08-22	checked all the electrical terminals of main supply	T. Dill	cl	working
21-08-22	checked temperature Controller Sensor for proper working	T. Dill	cl	working
28-08-22	checked heater Connections checked wiring coils for any leakage	T. Dill	cl	working
08-09-22	checked heater Connections	T. Dill	cl	working
12-09-22	Maintenance was done	T. Dill	cl	done
19-09-22	checked temperature Controller Sensor for proper working	T. Dill	cl	working
26-09-22	Dust cleaned	T. Dill	cl	done
01-10-22	checked all the electrical terminals of main supply	T. Dill	cl	working
13-10-22	checked heater Connections checked wiring coils for any leakage	T. Dill	cl	working
21-10-22	checked temperature Controller Sensor for proper working	T. Dill	cl	working
29-10-22	checked all the electrical terminals of main supply	T. Dill	cl	working
02-11-22	checked temperature Controller Sensor for proper working	T. Dill	cl	working
10-11-22	Maintenance was done	T. Dill	cl	done
19-11-22	Dust was cleaned	T. Dill	cl	done
25-11-22	checked heater Connections	T. Dill	cl	working
03-12-22	checked all the electrical terminals of main supply	T. Dill	cl	working
11-12-22	Maintenance was done	T. Dill	cl	done
19-12-22	checked temperature Controller Sensor for proper working	T. Dill	cl	working

PRINCIPAL

Aditya Pharmacy College

SURAMPALAM-533

